

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

APR 23 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130194</u> ✓
Name <u>SGT JAMES MICHELS</u>		ID# <u>6502</u> Date <u>4-21-09</u> ✓
A	Agency <u>USBPD</u>	Phone # <u>852-6111</u>
Instrument Location <u>SQUAD ROOM - BARROW</u> ✓		
Alco S/N <u>X172971</u> ✓		Target Value <u>.081</u> ✓ High Pressure <u>375</u>
B	Alco Test Values 1 st Alco <u>.084</u> ✓ 2 nd Alco <u>.084</u> ✓	
Signature <u>SGT JAMES MICHELS</u> ✓		<u>5/13/09</u>


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

(CONTINUED FROM FRONT PAGE)

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

~~DIAGNOSTIC CHECK~~

```

"##123456789:;<=>@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ_`abcde fghijklmno
pqrstuvwxyz{|}~"

```